



Certificate of Quality

Catalog Number 226-119

Sorbent: Silica Gel coated with
2,4-Dinitrophenylhydrazine
Tube Holder: Type B
Storage: Store in freezer before use. After
sampling, refrigerated storage and
shipping to the laboratory is required.

**SKC is an ISO 9001 registered
manufacturing company.**

The SKC Certificate of Quality assures that these sorbent tubes have been manufactured to meet or exceed the requirements for sorbent purity, sample collection and retention, storage stability, and desorption efficiency for the following chemical hazards and analytical methods published by NIOSH, OSHA, EPA or ASTM:

Acetone	ASTM	D5197
Aldehydes	ASTM	D5197
Aldehydes	EPA	TO-11A
Formaldehyde	ASTM	D5197
Formaldehyde	EPA	IP-6A
Formaldehyde	EPA	TO-11A
Formaldehyde	NIOSH	2016
Glutaraldehyde	NIOSH	2532
Hexanal	ASTM	D5197

For compliance sampling use these tubes with an accredited sampling method. Such a method does not purport to address all problems, if any, associated with their use. It is the user's responsibility, employing a suitable method, to establish appropriate safety and health practices and to determine the applicability of regulatory limitations before use. The user should adjust the sampling procedures to meet specific conditions and should test tubes to ensure that the desired results will be obtained.

Buyer shall not be entitled to recover from SKC Inc. any consequential damages, damages to property, damages for loss of use, loss of time, loss of profits, or income or other incidental damages. Nor shall buyer be entitled to recover from SKC Inc. any consequential damages resulting from defect of the instrument including, but not limited to, any recovery under section 402A of the Restatement, Second of Torts.



Operating Instructions

863 Valley View Road, Eighty Four PA 15330 USA
Tel: 724-941-9701 Fax: 724-941-1369 e-mail: skctech@skcinc.com

Sorbent Sample Tube Catalog No. 226-119

Description

Sorbent sample tube 226-119 is a two-section tube containing 300 and 150 mg of DNPH-coated silica gel, respectively. The second section acts as a backup section to detect sample breakthrough.

Sampling

The suggested active sampling flow rate is 200 ml/min within a flow range of 100 to 1000 ml/min. *Typical personal sampling pumps may not achieve a flow rate above 600 ml/min with this tube.* For additional sampling parameters, consult the method used. Sampling parameters may vary with ambient concentrations of analyte.

Interferences

Ozone interference is associated with DNPH-coated adsorbent tubes. Sorbent sample tube Cat. No. 226-120 (a 226-119 tube with built-in ozone scrubber) may be used in atmospheres where ozone is present.

Sorbent sample tube 226-119 is known to contain the hydrazone derivative of acetone, which may cause interference in the analysis for acrolein. Therefore, SKC does not recommend this tube for acrolein unless the preferred analytical technique resolves this interference.

Preparing Formaldehyde Samples for Analysis

1. Place each sorbent section into separate 3-dram glass vials.
2. Add 3 ml* of acetonitrile to each vial. Cap each vial.
3. Shake each vial periodically over a 30-minute period.

Analyzing Formaldehyde Samples

1. Analyze sample extracts by HPLC with UV detection at 365 nm.
2. Several tubes should be analyzed from the lot to determine background.
3. The background must be subtracted from the sample result.

* 3 ml is the minimum volume needed to achieve good recoveries. Larger extraction volumes may be used, but will result in higher limits of quantitation.