



863 Valley View Rd, Eighty Four, PA 15330
724-941-9701
www.skcinc.com

Surface SWYPE for Lithium Cat. No. 769-1019 Operating Instructions

Limited shelf-life: Check expiration date on packaging.

1. Lightly spray the area or item (e.g., workbench, tool, control knob) with Cleaning/Developing Solution for Lithium Cat. No. 769-1056.
2. Wait at least 30 seconds and then wipe with a Surface SWYPE pad.
3. Allow 3 to 5 minutes for the color to develop on the pad. Lithium will produce a bright pink color.

Notes: Read results only within the test window of between 3 and 5 minutes.

Known interference: Nickel. The presence of nickel can be confirmed with a Nickel Surface SWYPE Cat. No. 769-1028 and Developer Cat. No. 769-1068.

SKC Limited Warranty and Return Policy

SKC products are subject to the SKC Limited Warranty and Return Policy, which provides SKC's sole liability and the buyer's exclusive remedy. To view the complete SKC Limited Warranty and Return Policy, go to www.skcinc.com/warranty-information.



Distributed by
Air-Met Scientific
☎ 1800 000 744
✉ sales@airmet.com.au
🌐 www.airmet.com.au

Form 37570 Rev 0