

 Distributed by:

 Air-Met Scientific Pty Ltd

 Air-Met Sales/Service

 P: 1800 000 744

 P: 1800 000 744

 E: sales@airmet.com.au

 W: www.airmet.com.au

 W: www.airmet.com.au

Datasheet



Probe HFD-0191

# Isotropic measurement of magnetic fields from 27 MHz to 1 GHz

using the Field Meter FieldMan®

The probe detects magnetic fields from 27 MHz up to 1 GHz. This wide frequency range for a magnetic field probe covers all the major areas of electromagnetic field exposure that can occur in radio and TV broadcasting, telecommunications, and in high frequency industrial applications. The probe is suitable for the detection of human exposure limits in the general public and occupational environments.

The probe's interface digitally transmits the measurement data to the base unit, which has no individual influence on the measured values and therefore does not need to be calibrated. Probe calibration is carried out at several frequencies. The calibration data is stored in the probe and is automatically taken into account during the measurement. If the frequency of the predominating field strength is known, a correction factor can additionally be applied to increase the measurement accuracy.

- > Isotropic (non-directional) measurement
- > High dynamic range of 60 dB
- > Wide True RMS range up to 1 A/m
- > Digital probe interface no more meter calibration
- Self-test of the probe interface with integrated sensor function test
- Automatic offset correction, no zero adjustment required
- > Wide temperature range from -20 °C to +50 °C
- > High Immunity at 50/60 Hz
- > Factory calibration up to 1 GHz





# Specifications <sup>1</sup>

Product Features		
Frequency range <sup>2</sup>	27 MHz to 1 GHz, magnetic (H-)field	
Type of frequency response	Flat	
Measurement range (nom.)	0.016 to 16 A/m (CW) 0.016 to 1 A/m (True RMS)	9.6 μW/cm² to 10 W/cm² (CW) 9.6 μW/cm² to 38 mW/cm² (True RMS)
Dynamic range (nom.)	60 dB	
CW damage level (nom.)	20 A/m	15 W/cm <sup>2</sup>
Peak damage level (nom.) <sup>3</sup>	200 A/m	1.5 kW/cm <sup>2</sup>
Sensor type	Diode based system	
Directivity	Isotropic (Tri-axial)	
Spatial assessment	3 separate axes	
Sampling rate / integration time (nom.)	5 Hz / 270 ms	
Temperature sensors	Integrated sensors for displaying the ambient temperature and for automatic offset compensation	
Self-test	Interface function test and sensor test for interruption of diodes	

Uncertainty	ertainty		
Flatness of frequency response <sup>4, 5</sup> Calibration uncertainty not included	±0.7 dB (50 to 80 MHz) ±0.5 dB (>80 to 250 MHz) ±0.8 dB (>250 to 1000 MHz)		
Linearity deviation (nom.) Referred to 2 mW/cm² (0.23 A/m)	+2/-3 dB (0.022 to 0.05 A/m) ±1 dB (0.05 to 0.1 A/m) ±0.5 dB (0.1 to 3 A/m) ±1 dB (3 to 16 A/m)	+2/-3 dB (18 to 100 μW/cm²) ±1 dB (100 to 380 μW/cm²) ±0.5 dB (0.38 to 340 mW/cm²) ±1 dB (0.34 to 10 W/cm²)	
Isotropic deviation <sup>5</sup>	±1 dB +0.1/ -0,2 dB (0 °C to 50 °C, related to 23 °C) +0.1/ -0.5 dB (-20 °C to +50 °C, related to 23 °C)		
Temperature response (nom.) Referred to 2 mW/cm² (0.23 A/m)			

General Specification	neral Specifications	
Calibration		Factory calibration according to ISO 17025 and certified according to ISO 9001
Recommended calibration interval		24 months
Operating temperature		-20 °C to +50 °C
Humidity		< 29 g/m³ (< 93 % RH at +30 °C), non-condensing
Ingress protection		IP54 (probe screwed on)
	Storage	1K5 (IEC 60721-3) -40 °C to +70 °C
Climatic conditions	Transport	2K4 (IEC 60721-3) -40 °C to +70 °C
	Operating	7K2 (IEC 60721-3) extended to -20 °C to +50 °C
Size		308 mm x 66 mm Ø
Weight		< 100 g
Country of origin		Germany

<sup>&</sup>lt;sup>1</sup> Unless otherwise noted specifications apply at reference condition: device in far-field of source, ambient temperature 23±3 °C, relative air humidity 25% to 75%, sinusoidal signal, probe sampling rate 5 Hz.

 <sup>&</sup>lt;sup>2</sup> Cutoff frequency at typ. -3 dB.

<sup>&</sup>lt;sup>3</sup> Pulse length 1µsec, duty cycle 1:100.

<sup>&</sup>lt;sup>4</sup> Frequency response can be compensated for by the use of correction factors stored in the probe memory.

 <sup>&</sup>lt;sup>5</sup> Results are calculated from the maximum and minimum response obtained during the full revolution about the stem of the probe, oriented 54.7° to the electric field vector.



## **Definitions and Conditions**

#### Conditions

Unless otherwise noted, specifications apply after 30 minutes warm-up time within the specified environmental conditions. The product is within the recommended calibration cycle.

#### **Specifications with limits**

These describe product performance for the given parameter covered by warranty. Specifications with limits (shown as  $<, \leq, >, \geq, \pm, \max, \min$ .) apply under the given conditions for the product and are tested during production, considering measurement uncertainty.

#### **Specifications without limits**

These describe product performance for the given parameter covered by warranty. Specifications without limits represent values with negligible deviations, which are ensured by design (e.g. dimensions or resolution of a setting parameter).

### Typical values (typ.)

These characterize product performance for the given parameter that is not covered by warranty. When stated as a range or as a limit (shown as <,  $\leq$ , >,  $\geq$ ,  $\pm$ , max., min.), they represent the performance met by approximately 80% of the instruments. Otherwise, they represent the mean value. The measurement uncertainty is not taken into account.

#### Nominal values (nom.)

These characterize expected product performance for the given parameter that is not covered by warranty. Nominal values are verified during product development but are not tested during production.

#### Uncertainties

These characterize the dispersion of the values attributed to the measurands with an estimated confidence level of approximately 95%. Uncertainty is stated as the standard uncertainty multiplied by the coverage factor k=2 based on the normal distribution. The evaluation has been carried out in accordance with the rules of the "Guide to the Expression of Uncertainty in Measurement" (GUM).

## **Ordering Information**

Digital Broadband Probe	Part number
Probe HFD-0191, H-Field, 27 MHz–1 GHz	2462/06
Optional Accessories	Part number
Cable, Digital Probe Extension, 2m <sup>6</sup>	2460/90.02

<sup>6</sup> The device specifications apply without an extension cable.

## Narda Safety Test Solutions GmbH

Sandwiesenstrasse 7 72793 Pfullingen, Germany Phone +49 7121 97 32 0 info@narda-sts.com

www.narda-sts.com

#### Narda Safety Test Solutions North America Representative Office 435 Moreland Road Hauppauge, NY11788, USA Phone +1 631 231 1700 info@narda-sts.com

Narda Safety Test Solutions S.r.l. Via Benessea 29/B 17035 Cisano sul Neva, Italy Phone +39 0182 58641 nardait.support@narda-sts.it

#### Narda Safety Test Solutions GmbH Beijing Representative Office Xiyuan Hotel, No. 1 Sanlihe Road, Haidian 100044 Beijing, China Phone +86 10 6830 5870 support@narda-sts.cn

<sup>®</sup> Names and Logo are registered trademarks of Narda Safety Test Solutions GmbH - Trade names are trademarks of the owners.